

AUG 24 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u> ✓	
Name <u>Garrett Willis</u>		ID# <u>5560</u>	Date <u>8/20/09</u> ✓
A	Agency <u>AST</u>	Phone # <u>907-262-4453</u>	
Instrument Location <u>Cooper Landing</u> ✓			
Alco S/N <u>57345</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>1,000</u>
B	Alco Test Values		
	<u>.080</u> ✓ 1 st Alco	<u>.080</u> ✓ 2 nd Alco	
Signature <u>Garrett Willis</u> ✓		CAB 9/1/09	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

AUGUST 28, 2009

OPERATOR'S NAME:
WILLIS GARRETT/R

OPERATOR'S NUMBER: 5560

SUBJECT'S LAST NAME:
VOC

SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC

DEPT/AGENCY: C008

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .002

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.64 in

ALCO TARGET	.001	12:34
BLANK TEST	.000	12:35
INTERNAL STANDARD	VERIFIED	12:35
ALCO TV 29.65 in	.000	12:35
BLANK TEST	.000	12:36
STANDARD SAMPLE	.000	12:36
BLANK TEST	.000	12:37
ALCO TV 29.65 in	.000	12:37
BLANK TEST	.000	12:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130196

AUGUST 28, 2009
TIME 12:31

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.65 in

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF G
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi jk lmnop
qrstuvwxy{1}~